



For Professional Use Only

AmpliSens[®] *Trichomonas vaginalis*-FRT

PCR kit

Instruction Manual

AmpliSens[®]



Ecoli s.r.o., Studenohorska 12
841 03 Bratislava 47
Slovak Republic
Tel.: +421 2 6478 9336
Fax: +421 2 6478 9040



Federal Budget Institute of
Science "Central Research
Institute for Epidemiology"
3A Novogireevskaya Street
Moscow 111123 Russia

TABLE OF CONTENTS

1. INTENDED USE	3
2. PRINCIPLE OF PCR DETECTION.....	3
3. CONTENT	3
4. ADDITIONAL REQUIREMENTS	4
5. GENERAL PRECAUTIONS.....	5
6. SAMPLING AND HANDLING	6
7. WORKING CONDITIONS.....	6
8. PROTOCOL.....	6
9. DATA ANALYSIS.....	8
10. TROUBLESHOOTING.....	9
11. TRANSPORTATION.....	9
12. STABILITY AND STORAGE.....	10
13. SPECIFICATIONS.....	10
14. REFERENCES	11
15. QUALITY CONTROL.....	11
16. KEY TO SYMBOLS USED	12

1. INTENDED USE

AmpliSens[®] *Trichomonas vaginalis*-FRT PCR kit is an *in vitro* nucleic acid amplification test for qualitative detection of *Trichomonas vaginalis* DNA in the clinical materials (urogenital swabs, urine, and prostate gland secretion) by using real-time hybridization-fluorescence detection.



The results of PCR analysis are taken into account in complex diagnostics of disease.

2. PRINCIPLE OF PCR DETECTION

Trichomonas vaginalis detection by the polymerase chain reaction (PCR) is based on the amplification of pathogen genome specific region using special primers. In real-time PCR the amplified product is detected by using fluorescent dyes. These dyes are linked to oligonucleotide probes which bind specifically to the amplified product. Real-time monitoring of fluorescence intensities during the real-time PCR allows the detection of accumulating product without re-opening the reaction tubes after the PCR run. **AmpliSens[®] *Trichomonas vaginalis*-FRT** PCR kit is a qualitative test that contains the Internal Control (IC), which must be used in the extraction procedure in order to control the extraction process of each individual sample and to identify possible reaction inhibition. **AmpliSens[®] *Trichomonas vaginalis*-FRT** PCR kit uses “hot-start”, which greatly reduces the frequency of nonspecifically primed reactions. “Hot-start” is guaranteed by separation of nucleotides and Taq-polymerase by using a wax layer or a chemically modified polymerase (TaqF). Wax melts and reaction components mix only at 95 °C. Chemically modified polymerase (TaqF) is activated by heating at 95 °C for 15 min.

3. CONTENT

AmpliSens[®] *Trichomonas vaginalis*-FRT PCR kit is produced in 3 forms:

AmpliSens[®] *Trichomonas vaginalis*-FRT PCR kit variant FRT (for use with RG),

REF R-B6(RG)-CE.

AmpliSens[®] *Trichomonas vaginalis*-FRT PCR kit variant FRT (for use with iQ),

REF R-B6(iQ)-CE.

AmpliSens[®] *Trichomonas vaginalis*-FRT PCR kit variant FRT-100 F,

REF R-B6-F(RG,iQ)-CE.

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit variant FRT includes:

Reagent	Description	Volume (ml)	Amount
PCR-mix-1-FL <i>Trichomonas vaginalis</i> ready-to-use-single-dose-test-tubes (under wax)	colorless clear liquid	0.01	110 tubes of 0.2 ml
PCR-mix-2-FL-red	red clear liquid	1.1	1 tube
Positive Control complex (C+)	colorless clear liquid	0.2	1 tube
DNA-buffer	colorless clear liquid	0.5	1 tube
Negative Control (C-)*	colorless clear liquid	1.2	1 tube
Internal Control-FL (IC)**	colorless clear liquid	1.0	1 tube

*must be used in the extraction procedure as Negative Control of Extraction.

** add 10 µl of Internal Control-FL during the DNA extraction directly to the sample/lysis mixture (see the DNA-sorb-AM **REF** K1-12-100-CE protocol).

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit is intended for 110 reactions (including controls).

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit variant FRT-100 F includes:

Reagent	Description	Volume (ml)	Amount
PCR-mix-1-FL <i>Trichomonas vaginalis</i>	colorless clear liquid	1.2	1 tube
PCR-mix-2-FRT	colorless clear liquid	0.3	2 tubes
Polymerase (TaqF)	colorless clear liquid	0.03	2 tubes
Positive Control complex (C+)	colorless clear liquid	0.2	1 tube
DNA-buffer	colorless clear liquid	0.5	1 tube
Negative Control (C-)*	colorless clear liquid	1.2	1 tube
Internal Control-FL (IC)**	colorless clear liquid	1.0	1 tube

*must be used in the extraction procedure as Negative Control of Extraction.

** add 10 µl of Internal Control-FL during the DNA extraction directly to the sample/lysis mixture (see the DNA-sorb-AM **REF** K1-12-100-CE protocol).

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit variant FRT-100 F is intended for 110 reactions (including controls).

4. ADDITIONAL REQUIREMENTS

- DNA extraction kit.

- Transport medium.
- Disposable powder-free gloves and laboratory coat.
- Pipettes (adjustable).
- Sterile pipette tips with aerosol barriers up to 200 µl.
- Tube racks.
- Vortex mixer.
- Desktop centrifuge with rotor for 2-ml reaction tubes.
- PCR box.
- Rotor-Gene 3000 or Rotor-Gene 6000 (Corbett Research, Australia) Instrument; iCycler iQ and iQ5 (Bio-Rad, USA) Instrument.
- Disposable polypropylene microtubes for PCR (0.2- or 0.1-ml; for example, Axygen, USA).
- Refrigerator for 2–8 °C.
- Deep-freezer for ≤ –16 °C.
- Waste bin for used tips.

5. GENERAL PRECAUTIONS

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol barriers and use new tip for every procedure.
- Store and handle amplicons away from all other reagents.
- Thaw all components thoroughly at room temperature before starting detection.
- When thawed, mix the components and centrifuge briefly.
- Use disposable gloves, laboratory coats, protect eyes while samples and reagents handling. Thoroughly wash hands afterward.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a kit after its expiration date.
- Dispose of all samples and unused reagents in compliance with local authorities' requirements.
- Samples should be considered potentially infectious and handled in a biological cabinet in accordance with appropriate biosafety practices.
- Clean and disinfect all sample or reagent spills using a disinfectant, such as 0.5% sodium hypochlorite or another suitable disinfectant.
- Avoid contact with the skin, eyes, and mucosa. If skin, eyes, or mucosa contact immediately flush with water and seek medical attention.

- Material Safety Data Sheets (MSDS) are available on request.
- Use of this product should be limited to personnel trained in the techniques of DNA amplification.
- The laboratory process must be one-directional, it should begin in the Extraction Area and then move to the Amplification and Detection Areas. Do not return samples, equipment and reagents to the area in which the previous step was performed.



Some components of this kit contain sodium azide as a preservative. Do not use metal tubing for reagent transfer.

6. SAMPLING AND HANDLING



Obtaining samples of biological materials for PCR-analysis, transportation, and storage are described in the manufacturer's handbook [1]. It is recommended that this handbook is read before starting work.

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit is intended to analyze DNA extracted with DNA extraction kits from:

- *urogenital swabs*;
- *urine (a sediment of the first portion of the morning specimen)*;
- *prostate gland secretion*.

7. WORKING CONDITIONS

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit should be used at 18–25 °C.

8. PROTOCOL

8.1. DNA extraction

It is recommended that the following nucleic acid extraction kits are used:

- DNA-sorb-AM, **REF** K1-12-100-CE.
- Other nucleic acid extraction kits recommended by CRIE.



Carry out the DNA extraction according to the instructions provided by the manufacturer.

8.2. Preparing PCR

8.2.1 Preparing tubes for PCR

Variant FRT

The total reaction volume is **30 µl**, the volume of DNA sample is **10 µl**.

1. Collect the required number of the tubes with **PCR-mix-1-FL *Trichomonas vaginalis*** and wax for amplification of DNA from clinical and control samples.

- Add **10 µl** of **PCR-mix-2-FL-red** to the surface of wax layer of each tube, so that it does not fall under the wax and mix with **PCR-mix-1-FL *Trichomonas vaginalis***.

Variant FRT-100 F

Total reaction volume is **25 µl**, the volume of DNA sample is **10 µl**.

- Thaw the **PCR-mix-2-FRT** tube. Vortex the tubes with **PCR-mix-1-FL *Trichomonas vaginalis***, **PCR-mix-2-FRT**, and **polymerase (TaqF)**, then centrifuge briefly.

Collect the required number of the tubes/strips for amplification of DNA obtained from clinical and control samples.

- For N reactions (including 2 controls) add to a new tube:

10*(N+1) µl of **PCR-mix-1-FL *Trichomonas vaginalis***;

5.0*(N+1) µl of **PCR-mix-2-FRT**;

0.5*(N+1) µl of **polymerase (TaqF)**.

Vortex the tube, then centrifuge briefly. Transfer **15 µl** of the prepared mixture to each tube.

Steps 3 and 4 are applied for both variants.

- Add **10 µl** of **DNA samples** obtained from clinical or control samples at the stage of DNA extraction into the prepared tubes using tips with aerosol barrier.

- Carry out control amplification reactions:

NCA -Add **10 µl** of **DNA-buffer** to the tube labeled NCA (Negative Control of Amplification).

C+ -Add **10 µl** of **Positive Control complex** (to the tube labeled C+ (Positive Control of Amplification)).

C- -Add **10 µl** of a sample extracted from the **Negative Control** to the tube labeled C- (Negative Control of Extraction).

8.2.2. Amplification

Program the real-time amplification instrument according to manufacturer's manual and Guidelines [2].

- Create a temperature profile on your instrument as follows:

Table 1

AmpliSens-1 program

Step	Rotor-type Instruments ¹			Plate-type Instruments ²		
	Temperature, °C	Time	Cycles	Temperature, °C	Time	Cycles
1	95	15 min	1	95	15 min	1
2	95	5 s	5	95	5 s	5
	60	20 s		60	20 s	

¹ For example, Rotor-Gene 3000, Rotor-Gene 6000, Rotor-Gene Q or equivalent.

² For example, iCycler, iQ5, Mx3000P, Mx3000, DT-96 or equivalent.

	72	15 s		72	15 s	
3	95	5 s	40	95	5 s	40
	60	20 s <i>fluorescent signal detection</i>		60	30 s <i>fluorescent signal detection</i>	
	72	15 s		72	15 s	

Fluorescence detection should be enabled in the channels designed for FAM and JOE fluorophores.

Adjust the fluorescence channel sensitivity according to *Important Product Information Bulletin* enclosed to the PCR kit.

9. DATA ANALYSIS

The fluorescent signal intensity is detected in two channels:

- The signal from the *Trichomonas vaginalis* DNA amplification product is detected in the FAM channel;
- The signal from the Internal Control amplification product is detected in the JOE channel.

Result interpretation

The results are interpreted with the Instrument software by the crossing (or not crossing) of the fluorescence curve with a threshold line and showed as presence (or absence) of Ct (cycle threshold) in the result grid.

Principle of interpretation:

- *Trichomonas vaginalis* DNA is **detected** in a sample if its Ct value is defined in the result grid in the FAM channel. Moreover, the fluorescence curve should cross the threshold line in the region of fluorescence exponential growth.
- *Trichomonas vaginalis* DNA is **not detected** in a sample if its Ct value is not defined in the result grid in the FAM channel (the fluorescence curve does not cross the threshold line), whereas the Ct value in the JOE channel is less than the specified boundary Ct value.
- The result is **invalid** if Ct of a sample in the FAM channel is absent whereas the Ct value in the JOE channel is either absent or greater than the specified boundary Ct value. It is necessary to repeat the PCR test for such a sample.



Boundary Ct values are specified in the *Important Product Information Bulletin* enclosed to the PCR kit.

Result of the analysis is considered reliable only if the results for both Positive and Negative Controls of amplification as well as Negative Control of extraction are correct (Table 2).

Table 2

Results for controls

Control	Stage for control	Ct value in channel		Interpretation
		FAM fluorophore	JOE fluorophore	
C-	DNA extraction	Neg	Pos (< boundary value)	OK
NCA	Amplification	Neg	Neg	OK
C+	Amplification	Pos (< boundary value)	Pos (< boundary value)	OK

10. TROUBLESHOOTING

Results of analysis are not taken into account in the following cases:

1. If the Ct value is absent in both JOE/Yellow/HEX and FAM/Green channels or the Ct value in the JOE/Yellow /HEX channel is higher than the specified boundary Ct value, PCR should be repeated. If the same result is obtained, the extraction stage for the sample should be repeated. If the IC signal of this sample was detected normally in any other PCR test, it is not necessary to repeat the extraction stage (for iCycler iQ or iQ5 instruments).
2. If a Ct value is detected for C- in the FAM/Green channel and/or for NCA in the FAM, JOE/Yellow /HEX channels in the results grid, this indicates contamination of reagents or samples. In such cases, the results of analysis must be considered as invalid. Test analysis must be repeated and measures to detect and eliminate the source of contamination must be taken.
3. If no signal is detected for the Positive Controls of amplification, it may suggest that the programming of the temperature profile of the used Instrument was incorrect, or that the configuration of the PCR reaction was incorrect, or that the storage conditions for kit components has not complied with the manufacturer's instruction, or that the reagent kit has expired. Programming of the used instrument, storage conditions, and the expiration date of the reagents should be checked, and then PCR should be repeated.
4. If a positive result (the fluorescence curve crosses the threshold line) is detected for a sample that has a fluorescence curve without the typical exponential growth phase (the curve is linear), this may suggest incorrect setting of the threshold line or incorrect calculation of baseline parameters. Such a result should not be considered as positive. Once the threshold line has been set correctly, PCR analysis of the sample should be repeated (if iCycler iQ or iQ5 instruments are used).

If you have any further questions or if encounter problems, please contact our Authorized representative in the European Community.

11. TRANSPORTATION

AmpliSens® *Trichomonas vaginalis*-FRT PCR kit should be transported at 2–8 °C for no longer than 5 days.

12. STABILITY AND STORAGE

All components of the **AmpliSens® *Trichomonas vaginalis*-FRT** PCR kit (except for Polymerase (TaqF) and PCR-mix-2-FRT) are to be stored at 2–8°C when not in use. They are stable until the expiration date on the label. The shelf life of opened reagents is the same as that of unopened reagents, unless otherwise stated.



Polymerase (TaqF) and PCR-mix-2-FRT are to be stored at temperature from minus 24 to minus 16 °C when not in use.



PCR-mix-1-FL *Trichomonas vaginalis* should be kept away from light.

13. SPECIFICATIONS

13.1. Sensitivity

The analytical sensitivity of **AmpliSens® *Trichomonas vaginalis*-FRT** PCR kit is specified in the table below.

Clinical material	Transport medium	DNA extraction kit	Analytical sensitivity, GE/ml*
Urogenital swabs	Transport Medium for Swabs (REF 956-CE, REF 987-CE) or Transport Medium with Mucolytic (REF 952-CE, REF 953-CE)	DNA-sorb-AM	5 x 10 ²
Urine (pretreatment is required)	–	DNA-sorb-AM	1 x 10 ³

* Genome equivalents (GE) of the microorganism per 1 ml of a clinical sample placed in the transport medium specified.

13.2. Specificity

The analytical specificity of **AmpliSens® *Trichomonas vaginalis*-FRT** PCR kit is ensured by selection of specific primers and probes as well as stringent reaction conditions. The primers and probes were checked for possible homologies to all sequences published in gene banks by sequence comparison analysis. There were not nonspecific test responses during examination of human DNA as well as a DNA panel of the following microorganisms: *Gardnerella vaginalis*, *Lactobacillus* spp., *Escherichia coli*, *Staphylococcus aureus*, *Streptococcus pyogenes*, *Streptococcus agalactiae*, *Candida albicans*, *Mycoplasma hominis*, *Ureaplasma urealyticum*, *Ureaplasma parvum*, *Neisseria flava*, *Neisseria subflava*, *Neisseria sicca*, *Neisseria mucosa*, *Neisseria gonorrhoeae*, *Chlamydia trachomatis*, *Treponema pallidum*, *Toxoplasma gondii*, HSV types 1 and 2, CMV, and HPV.

The clinical specificity of **AmpliSens® *Trichomonas vaginalis*-FRT** PCR kit was confirmed in laboratory clinical trials.














14. REFERENCES

1. Handbook “Sampling, Transportation, and Storage of Clinical Material for PCR Diagnostics”, developed by Federal Budget Institute of Science “Central Research Institute for Epidemiology” of Federal Service for Surveillance on Consumers’ Rights Protection and Human Well-Being, Moscow, 2008.
2. Guidelines “Real-Time PCR Detection of STIs and Other Reproductive Tract Infections” developed by Federal Budget Institute of Science “Central Research Institute for Epidemiology” of Federal Service for Surveillance on Consumers’ Rights Protection and Human Well-Being, Moscow.

15. QUALITY CONTROL

In accordance with Federal Budget Institute of Science “Central Research Institute for Epidemiology” ISO 13485-Certified Quality Management System, each lot of **AmpliSens® *Trichomonas vaginalis*-FRT** PCR kit has been tested against predetermined specifications to ensure consistent product quality.

16. KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Sufficient for
	<i>In vitro</i> diagnostic medical device		Expiration Date
	Version		Consult instructions for use
	Temperature limitation		Keep away from sunlight
	Manufacturer	NCA	Negative control of amplification
	Date of manufacture	C-	Negative control of extraction
	Authorised representative in the European Community	C+	Positive control of Amplification
FBIS CRIE	Federal Budget Institute of Science “Central Research Institute for Epidemiology”	IC	Internal control

List of Changes Made in the Instruction Manual

VER	Location of changes	Essence of changes
29.06.11 LA	Cover page, text	The name of Institute was changed to Federal Budget Institute of Science “Central Research Institute for Epidemiology”